

Tribological Behavior of Bone Against Calcium Titanate Coating in Simulated Body Fluid

Comportamiento tribológico de hueso contra recubrimiento de titanato de calcio en fluido corporal simulado

Esguerra-Arce Johanna

Escuela de Ingeniería de Materiales Universidad del Valle, Colombia E-mail: johanna.esguerra@correounivalle.edu.co

Aguilar-Castro Yesid

Escuela de Ingeniería de Materiales Universidad del Valle, Colombia E-mail: yesid.aguilar@correounivalle.edu.co

Aperador-Chaparro William

Facultad de Ingeniería Universidad Militar Nueva Granada, Bogotá, Colombia E-mail: post.ingenieria@unimilitar.edu.co

Ipaz-Cuastumal Leonid

Escuela de Ingeniería de Materiales Universidad del Valle, Colombia E-mail: leoipazc@gmail.com

Bolaños-Pantoja Gilberto

Laboratorio de Física de Bajas Temperaturas Universidad del Cauca, Colombia E-mail: gbolanos@unicauca.edu.co

Rincón-López Carlos Alberto

Laboratorio de Física de Bajas Temperaturas Universidad del Cauca, Colombia E-mail: crincon@unicauca.edu.co Information on the article: received: February 2014,

reevaluated: March 2014, accepted: March 2014

Abstract

Although calcium titanate has been proposed as a coating for biomedical applications, a characterization of tribological properties simulating human conditions has not been reported. In this work we studied friction and wear mechanism of calcium titanate coating growth onto AISI 304 steel (750 nm thickness) deposited by r.f. magnetron sputtering. It was found that the wear mechanisms of the system is bone adhesion to the coating without detachment of the coating, both dry and in Hank's solution, with a friction coefficient of 0.84 ± 0.13 and 0.65 ± 0.13 , respectively. The wear of the bone was more severe when using a simulated body fluid at 37° C in the pin on disk test.

Keywords:

- biomaterials
- calcium titanate coating
- tribology
- pin-on-disk
- · Hank's solution

Resumen

Aunque el titanato de calcio ha sido propuesto como recubrimiento para aplicaciones biomédicas, no se han reportado caracterizaciones tribológicas en condiciones que simulan las condiciones del interior del cuerpo humano. En este trabajo se evalúan propiedades de fricción y mecanismos de desgaste de estos recubrimientos (de 750 nm de espesor) depositados mediante r.f. magnetrón sputtering. Se encontró que el mecanismo de desgaste del sistema es adhesión de hueso al recubrimiento sin desprendimiento de la capa, tanto en seco como en solución de Hank, con coeficientes de fricción de 0.84 ± 0.13 y 0.65 ± 0.13 , respectivamente. El desgaste del hueso fue más severo cuando el ensayo de pin-disco se llevó a cabo en fluido corporal simulado.

Introduction

Hip prosthesis stems may be cemented or uncemented, the uncemented have four options for achieving adhesion with the surrounding tissue: by pressure, porous surface, screws or with coated surface. Joint prosthesis with hydroxyapatite coated stems have presented a big clinical success, as they increase the rate of osseointegration regarding uncemented prosthesis (Faig and Gil, 2008; Coathup *et al.*, 200; Rokkum and Reigstad, 1999). However, in the long run this coating dissolves and detaches exposing the base metal (Porter *et al.*, 2004), which can lead to the loosening of the prosthesis, pain in the patient, wear on the rest of the prosthesis or, in the worst case, to extensive destruction of bone tissue.

Although the calcium titanate has been studied extensively in regard to optical and luminescent properties (Moreira et al, 2009; Wang et al., 2009; Yang et al., 2009), an interface layer containing calcium titanate on biomedical titanium alloys coated with hydroxyapatite has been found (Ergun et al., 2003; Kokubo et al., 2003), so that some authors have reported to have obtained calcium titanate layers by various methods on titanium and titanium alloys, such as sol-gel (Holliday and Stanishevsky, 2004), magnetron sputtering (Asami et al., 2005), ion-beam assisted deposition (Ohtsu et al., 2006), pulsed laser deposition (Ohtsu et al., 2007), hydrothermal-electrochemical method (Wiff et al., 2007), thick alkalized calcium oxidation (Ohtsu et al., 2007), and hydrothermal treatment (Park et al., 2011) for biomedical applications.

Despite the fact that most commercial prostheses are made of titanium or titanium alloys, because the material replaced for bone is expected to have a modulus equivalent to that of bone (Geetha *et al.*, 2009), and due to their high cost and their poor tribological properties, stainless steels is still used in hip prosthesis and remains under investigation for biomedical purposes

Descriptores:

- biomateriales
- recubrimiento de titanato de calcio
- tribología
- pin-disco
- solución de Hank

(Lundin *et al.*, 2012; Nie, 2011; Barragán *et al.*, 2010). Although type 304 is seldom used in biomedical implants or devices and its sensitivity to any sign of corrosion is more detectable than in 316L (Tang *et al.*, 2006), it is used in implants such as brackets and screws. This steel also meets ASTM F138 y ASTM F139 (ASTM F138–08, 2008; ASTM F139–08, 2008) standards for steels used in biomedical applications and it is cheaper than AISI 316L which would represent a reduction in hip prosthesis cost. This is the reason for proposing here the evluation of tribological properties of AISI 304 stainless steel coated with calcium titanate for prosthesis stem applications.

It is known that due to the continuous micro-movements between the bone and the implant (Fu *et al.*, 1998) wear femoral stem occurs (Howell *et al.*, 1999), so it is important to know the tribological properties of the coating with respect to the bone, such as friction and wear in conditions which approach those of the human body. Therefore, this work aims to evaluate calcium titanate coated AISI 304 stainless steel by r.f. magnetron sputtering, evaluating mechanical properties of hardness and elastic modulus by nanoindentation; coefficient of friction and wear by pin-on-disc test in dry conditions at room temperature and in Hank's solution at 37 °C with an animal bone pin as a counterpart.

Materials and methods

The CaTiO₃ cathode used for the sputtering technique were from CaTiO₃ powder with a purification of 99.9% from Super Conductor Materials, Inc. Silicon (100) and 304 stainless steel disks (diameter = 1.27 cm and thickness = 3 mm) were used as substrates. The steel was polished with sandpaper passing numbers 200 to 2000, and finally with 1 and 0.3 μ m alumina solution. Prior to deposition of calcium titanate an approximately 200 nm

titanium layer was deposited onto steel substrate by magnetron sputtering, during the 3 hours at a temperature of 250°C, with a power density of the cathode fixed at 350 W.

Calcium titanate was deposited via magnetron sputtering in Ar atmosphere. Prior to the deposition, the deposition chamber was evacuated to a base pressure of 2×10^{-6} mbar and the target was sputter-cleaned for 20 min, before film growth. During the 4 hours deposition process at a temperature of 500°C, the sputtering power density deposition of the CaTiO₃ cathode was fixed at 2.47 W/cm², the pressure in the chamber was fixed at 2×10^{-2} mbar, and the target was fixed 4.3 cm under the substrate.

The crystal structure of the films was determined by using glancing incident X-ray diffraction (GIXRD) at 2° incidence angle with a RIGAKU (Dmax2100) difractrometer using Co K α radiation (λ = 1.78899 Å, 30 kV and 16 mA). The chemical composition of deposited films was performed by energy dispersive X-ray spectroscopy (EDX) in the scanning electron microscope (JEOL JSM-649 OLV SEM). The thickness of the films was measured by Scanning Electron Microscope (JEOL JSM-649 OLV). The mechanical analysis was performed via nanoindentation (ASTM E2546-07; 2007) by means of anUBIL - HYSITRON device and a diamond Berkovich tip. For each sample several series of ten indents were made and the results were averaged, since the maximal nanoindentation depth was about 10% of coating thickness.

grade its mechanical properties. Then the soft tissue was removed mechanically by hand and the bone was immersed in a bleach-water mixture (1:5 volume) for 53 hours for further machining to a spherical geometry of 6 mm diameter. For mechanical measurement via nanoindentation specimens were embedded in plastic (Lucite, Buehler) to protect them from damage and to provide a uniform format for preparation. The sample was polished using a sequence of abrasives silicon carbide grit paper ending with 0.3 micron alumina suspension, using water as a lubricant. Each sample was indented 24 times using a Berkovich pyramidal indenter, varying the maximum load applied from 479.6 to 5995.8 μ N (10 seconds of loading and 10 seconds of unloading).

Tribological test was carried out in dry and wet conditions (three bare steel samples and three coated samples in every condition). Dry tests were carried out at room temperature and relatively humidity of $65\% \pm 4$, wet tests, simulating body conditions (SBC), were performed using Hank's solution (Table 1) at $37^{\circ}C \pm 1$. The test is carried out with Hank's solution because this is, of all the available solutions, the one that most closely approximates to the composition of blood plasma (Zhanga *et al.*, 2013). The applied load was 3 N, the distance was 200 meters at 100 rpm (test duration: 1hour, 46 minutes). A detailed characterization of the worn surfaces was performed using a Scanning Electron Microscope (JEOL JSM-649 OLV) after cleaning with acetone and water at 50°C.

Taking into account that these coatings are proposed for use in hip prosthesis stems a pin-ondisk test (ASTM G99-05, 2010) with animal bone ball as a counterpart was used to measure the wear resistance and the friction coefficient using a MT60 tribometer from NANO-VEA[™]. The bone pin was prepared from a bovine metacarpal diaphysis, which was immersed in boiling water avoiding touching the bottom of the vessel to not de-

Table 1. Chemical composition of Hank's solution, concentration [g/L]

NaCl	KCl	MgSO ₄ .7H ₂ O	CaCl2.H ₂ O	Na_2HPO_4	$\rm KH_2PO_4$	NaHCO ₃	Glucosa
8.0	0.4	0.2	0.185	0.046	0.06	0.35	1.0

Table 2. Elemental chemical composition in calcium titanate coating





Results and discussion

Chemical composition, morphology and phase composition of the coatings

The EDX spectrum of elemental composition of the coating surface, Figure 1: a) revealed that the coating contains Ca, Ti and O atoms (Si belongs to the substrate) with a Ca-Ti ratio of 0.98. Table 2 shows the elemental chemical composition. As can be seen in b), the coating obtained is homogeneous without micro-pores or particles on the surface. Finally, a thickness of 750 nm is shown in a cross sectional view c) on calcium titanate coating.

The grazing incident XRD pattern of CaTiO₃ film is shown in Figure 2. This indicates that the coating exhibits polycrystalline structure with a (022) preferential orientation corresponding to Pbnm orthorhombic Ca-TiO₃ phase at 33.293°. The shift toward high angles (0.69 grades) at positions 33,293° and 40,673° of JCPDF 01-088-0790 card is in relation with compressive residual stress characteristics. The positions 52.827°, 52.931°, 64.747° and 64.866° are in agreement with JCPDF 01-082-0231 card which also corresponds to an orthorhombic Pbnm calcium titanate. The peak located at 72.779° is in agreement with JCPDF 00-043-0226 card, corresponding to a cubic *Pm-3m* calcium titanate, which is a phase that normally occur above 1300°C (Roushown and Masatomo, 2005), and it is known that metastable phases can be formed in as-deposited thin films by physical vapor deposition whereas they are hardly seen in bulk counterparts (Krzanowski, 2004). Finally, anatase phase with (101) and (224) reflections of TiO₂ was observed, in agreement with JCPDF 01-083-2243 card; and was also found a coincidence with 01-074-1226 card, in 64.647° position, which correspond to (311) reflection of CaO; that is because the deposition process is assumed to atomically disassemble the compound AXY, directing A, X and Y atoms toward the



Figure 2. GI-XRD patterns of the calcium titanate coating

substrate, where they can be reassembled in multiple phases (Krzanowski, 2004), it means that a part of the material reaching the substrate is suffering the next change:

$$CaTiO_3 \rightarrow TiO_2 + CaO$$

The crystallinity of the coatings indicates that heating the substrate at 500°C is appropriate for obtaining calcium titanate crystalline coatings. This is better than that reported by Naofumi Ohtsu *et al.* (2007) who obtained crystalline films at 600°C.

Mechanical properties

Hardness

Table 3 shows materials hardness and elastic modulus measured by nanoindentation. In all three cases the averaged uniaxial Young's moduli and hardness (in the axial and transverse directions for bone) values were computed from the experimental curves using the Oliver-Pharr theory (Oliver and Pharr, 1992). The mechanical differences in bone values (longitudinal and transversal directions) are due to structural anisotropy and are also linked to a variation in chemical composition (Rho *et al.*, 1998).

Tribological properties

In part a) of Figure 3 it can be seen the wear mechanisms on bare steel and the spherical counterpart bone. There are two wear mechanisms: abrasive wear and adhesive wear. Second zoom shows plow lines indicating abrasive wear likely because of some steel particle came off due to fatigue and strain hardened generated abrasion to disk. The zoom shows bone adhesion to the disk; it is also observed that adhered material fractures because of fatigue forming wear debris acting as a third body. In part b) we observe wear mechanisms of coated steel. Abrasive wear is not observed and EDX probe revealed no bare steel, so it follows that the coating maintained its integrity after testing. The only wear mechanism observed is adhesive and fatigue fracture of bonded bone forming wear particles. Figure 4 shows the wear mechanisms of bare steel and coated steel in Hank's solution. It can be seen that the only wear mechanisms both cases is bone adhesion, there is no abrasive wear in neither case, there is no detachment of the coating, and there is salts precipitations.

Figure 5 parts a) and b) show bone pin acting as counterpart of bare steel and coated steel, respectively,

DOI: https://doi.org/10.1016/j.riit.2015.03.011

Table 3. Hardness and elastic modulus of bare steel, calcium titanate coated steel and bone

Duonontry	Bare steel	CT coating	Bone		
Toperty			Longitudinal	Transversal	
Hardness [GPa]	5.3 ± 0.7	6.3 ± 0.1	0.57 ± 0.05	0.67 ± 0.05	
Elastic Modulus [GPa]	234 ± 12	133.69 ± 10	16.27 ± 1.03	21.06 ± 1.99	



Figure 3. Wear mechanism in a) the AISI 304 steel and, b) calcium titanate coated AISI 304 steel in pin on disc test using a bovine pin as a counterpart in *dry conditions*



Figure 4. Wear mechanism in a) the AISI 304 steel and, b) calcium titanate coated AISI 304 steel in pin on disc test using a bovine pin as a counterpart in Hank's solution at 37 $^\circ C$



Figure 5. Wear in pin acting like counterpart to a) bare steel in dry conditions, b) coated steel dry conditions, c) wear zone in pin, d) bare steel in Hank's solution, e) coated steel in Hank's solution f) debris from dry test

in dry test. Part c) presents how bone layers are formed due to fatigue, this layers subsequently will be bonded to the steel samples. Parts d) and e) show bone pin acting as counterpart of bare steel and coated steel, respectively, in Hank's solution test. Part f) shows the geometry of wear debris from dry test, which exhibit brittle fracture and an average size of 8 µm larger ones.

Figure 6 shows friction coefficient in function of sliding distance in pin-on-disk test. In part a) stages 1 and 2 refer to asperities deformation and contaminant removal mentioned by Holmberg and Matthews, (2009) appear in small proportion in bare steel and are almost nonexistent in coated steel due to their low initial roughness and high cleanness. Stage 3 where the friction coefficient increases due to the rapid increase of wear particles trapped between the sliding surfaces is larger in bare steel than in coated steel. This could be because the coating removes more bone wear debris due to its higher hardness, as can be seen in Table 3, which rapidly increases the friction coefficient. Relative to stage 4, one of the bare steel samples does show a short interval with a friction coefficient of about 1.7 before reaching the steady state, which may be related to greater wear bone, 1.59×10^2 mm³, compared with the wear in the bone provided because of the other two samples, 0.79×10^2 mm³ and 0.76×10^2 mm³. The reason for this differences in bone wear is the bone anisotropy: its hardness in the longitudinal direction is different in the transversal direction (Rho 1998); as this orientation could not be controlled in the pins preparation of the current investigation, these may exhibit different hardness according to the area that is in contact with the disk, in this case the hardness of the pin which acted as counter-part of the steel sample in question is 50.83 ± 3,59 HV, compared with the hardness of the other two samples, 54.23 ± 2.66 HV and 69.04 ± 1.94 HV, respectively. It can also make a correlation between bone pin wear and dynamic friction coefficient of coated steel: high to low friction coefficient pin wear were 1.04×10^2 mm³, 2.53×10^2 mm³ and 2.85×10^2 mm³, whose hardness were respectively



Figure 6. Friction coefficient for bare steel and calcium titanate coated steel in pin-on-disk test in a) dry conditions and b) simulated body conditions



Figure 7. a) Volume loss of bone ball counterpart in pin on disk test in dry and SBF conditions, b) bone hardness in function of immersion time in Hank's solution

59.77 ± 2.13 HV, 54.21 ± 3.28 HV and 43.74 ± 1.97 HV. Namely in dry conditions, volume loss of bone decreases with increasing bone hardness. At steady state it can be seen that friction coefficient of coated steel (0.84 ± 13) is on average greater than the coefficient of friction of bare steel (0.77 ± 0.11), which is related, as shown in Figure 7, with bone volume loss, $0.021 \text{ mm}^3 \pm 0.009 \text{ mm}^3$ for coated steel and $0.010 \text{ mm}^3 \pm 0.004 \text{ mm}^3$ for bare steel. It means a higher friction coefficient goes together with a higher volume loss.

Under conditions which simulate the human environment (Hank's solution at 37°C), on the contrary, part b) of Figure 6 shows that at steady state friction coefficient of bare steel (0.89 ± 0.12) is on average greater than the friction coefficient of coated steel (0.65 ± 0.12) related with bone volume loss of $0.051 \text{ mm}^3 \pm 0,009 \text{ mm}^3$ and 2.82 mm³ ± 0.58 mm³, respectively, as can be seeing in Figure 7. It means, in Hank's solution, a higher friction coefficient goes together with a lower volume loss.

Part a) of Figure 7 shows the loss in volume of the bone spherical counterpart in both test conditions. It can be observed that in dry conditions and in Hank's solution at 37°C, coated steel samples cause more wear in bone pin because of its greater hardness compared to bare steel. Analyzing each material is observed that wear was more severe when test was performed in Hank's solution at 37°C than when it was in dry conditions. This can be attributed to a decrease in mechanical properties of the bone immerse in a liquid environment which simulates blood plasma, as evidenced in part b) of Figure 7, which shows that bone loses hardness if it is immersed in the fluid, and lower hardness leads to increased wear.

Conclusions

- As depositing calcium titanate by magnetron sputtering onto AISI 304 stainless steel using a CaTiO₃ as cathode obtained by powder technology, a coating was obtained consisting of *Pbnm orthorrombic calcium titanate*, *Pm*–3*m cubic calcium titanate*, *titanium oxide* (anatase) and calcium oxide. The thickness of the coating was about 750 nm, with a hardness of 6.3 ± 0.1 and an elastic modulus of 133.69 ± 10.
- When in contact, polished steel AISI 304 with spherical bovine bone in pin-on-disc tested with a load of 3 N, the wear mechanisms that occur are pin volume loss, bone adhesion to the steel and steel abrasion. Under the same conditions but with calcium titanate coated steel as a flat counter-part, wear bone also occurs and bone adhesion to the disc without abrasion or detachment of the coating. The friction coefficient in each case was 0.77 ± 0.11 and 0.84 ± 0.13 .
- In Hank's solution, when in contact with both AISI 304 and calcium titanate coated, AISI 304 with spherical bovine bone in pin-on-disc tested with a load of 3 N, the wear mechanisms that occur are pin volume loss and bone adhesion to the disks. Friction coefficients were 0.89 ± 0.12 and 0.65 ± 0.12.
- It can be observed that in case of mutual movement between bone and steel in bio-

Esguerra-Arce Johanna, Aguilar-Castro Yesid, Aperador-Chaparro William, Ipaz-Cuastumal Leonid, Bolaños-Pantoja Gilberto, Rincón-López Carlos Alberto

logical medium, the bone damage is more severe if steel is coated with calcium titanate. Because of that it is important to measure the adherence of the coating to the bone, in order to analyze its viability as a coating in hip stems.

References

- Asami K., Ohtsu N., Saito K., Hanawa T. CaTiO3 films sputterdeposited under simultaneous Ti-ion implantation on Tisubstrate. *Surface & Coatings Technology*, volume 200, 2005: 1005-1008.
- ASTM E2546–07: Standard Practice for Instrumented Indentation Testing, West Conshohocken, PA, American Society for Testing and Materials, 2007.
- ASTM G99–05 (Reapproved 2010): Standard Test Method for Wear Testing with a Pin-on-Disk Apparatus, West Conshohocken, PA, American Society for Testing and Materials, 2010.
- ASTM F138 08: Standard Specification for Wrought 18Chromium-14Nickel-2.5Molybdenum Stainless Steel Bar and Wire for Surgical Implants (UNS S31673), West Conshohocken, PA, American Society for Testing and Materials, 2008.
- ASTM F139 08: Standard Specification for Wrought 18Chromium-14Nickel-2.5Molybdenum Stainless Steel Sheet and Strip for Surgical Implants (UNS S31673), West Conshohocken, PA, American Society for Testing and Materials, 2008.
- Barragán F., Guardián R., Menchaca C., Rosales I., Uruchurtu J. Electrochemical Corrosion of Hot Pressing Titanium Coated Steels for Biomaterial Applications. *International Journal of Electrochemical Science*, volume 5, 2010: 1799-1809.
- Coathup M. J., Blunn G. W., Flynn N., Williams C., Thomas N.P. A comparison of bone remodeling around hydroxyapatite-coated, porous-coated and grit-blasted hip replacements retrieved at post-mortem. *The Journal of Bone and Joint Surgery [Br]*, volume 82-B, 2000: 118-123.
- Ergun C., Doremus R., Lanford W.J. Hydroxyapatite and titanium-interfacial reactions. *Journal of Biomedicals Materials Research*, Part A, volume 65, 2003: 336.
- Faig-Martí J. and Gil-Mur F.J. Los recubrimientos de hidroxiapatita en las prótesis articulares. *Revista española de cirugía ortopédica y traumatología*, volume 52, 2008: 113-120.
- Fu Y., Batchelor A.W., Wang Y., Khor K.A. Fretting wear behaviors of thermal sprayed hydroxyapatite (HA) coating under unlubricated conditions. *Wear*, volume 217, 1998: 32-139.
- Geetha M., Singh A.K., Asokamani R., Gogia A.K. Ti based biomaterials, the ultimate choice for orthopaedic implants –A review. *Progress in Materials Science*, volume 54, 2009: 397-425.
- Holliday S. and Andrei S. Crystallization of CaTiO3 by sol-gel synthesis and rapid thermal processing. *Surface & Coatings Technology*, volume 188 and 189, 2004: 741-744.
- Holmberg K. and Matthews A. *Coatings tribology: Properties, mechanisms, techniques and applications in surface engineering,* volu-

me 56, 2nd ed., United Kingdom, Tribology and Interface Engineering Series, 2009, 48p.

- Howell J.R., Blunt L.A., Ling R.S.M. An analysis of fretting damage seen on explanted femoral stems. *The journal of bone and joint surgical*, volume 81B (Suppl. 3), 1999: 163.
- Kokubo T., Kim H.M., Kawashita M. Novel bioactive materials with different mechanical properties. Biomaterials, 24: 2161, 2003.
- Krzanowski J.E. Phase formation and phase separation in multiphase thin film hard coatings. *Surface & Coatings Technology*, volume 188 and 189, 2004: 376-383.
- Lundin M., Hedberg Y., Jiang T., Herting G., Wang X., Thormann E., Blomberg E., Odnevall I.W. Adsorption and protein-induced metal release from chromium metal and stainless Steel. *Journal of Colloid and Interface Science*, volume 366, 2012: 155-164.
- Moreira M.L., Paris E.C., Do Nascimento G.S., Longo V. M., Sambrano J.R., Mastelaro-Valmor R., Bernardi Maria I.B, Varela Juan Andrés José A., Longo Elson. Structural and optical properties of CaTiO3 perovskite-based materials obtained by microwave-assisted hydrothermal synthesis: An experimental and theoretical insight. *Acta Materialia*, volume 57, August, 2009: 5174-5185.
- Nie F.L., Wang S.G., Wang Y.B., Wei S.C., Zheng Y.F. Comparative study on corrosion resistance and in vitro biocompatibility of bulk nanocrystalline and microcrystalline biomedical 304 stainless Steel. *Dental materials*, volume 27, 2011: 677-683.
- Ohtsu N., Kesami S., Katsuhiko A., Takao H. Characterization of CaTiO3 thin film prepared by ion-beam assisted deposition. *Surface & Coatings Technology*, volume 200, 2006: 5455-5461.
- Ohtsu N., Akihiko I., Kesami S., Takao H. Characterization of calcium titanate thin films deposited on titanium with reactive sputtering and pulsed laser depositions. *Surface & Coatings Technology*, volume 201, 2007: 7686-7691.
- Ohtsu N., Chikage A., Tetsuya A., Satoshi S., Kazuaki W. Calciumhydroxide slurry processing for bioactive calcium-titanate coating on titanium. *Surface & Coatings Technology*, volume 202, 2008: 5110-5115.
- Oliver W.C., Pharr G.M. An improved technique for determining hardness and elastic modulus using load and displacement sensing indentation experiment. *Journal of Materials Research*, volume 7, 1992: 1564-1583.
- Park-JinW., Yusuke T., Chong-Soo L., Chan-Hee P., Youn-Jeong K., Je-Hee J., Dongwoo K., Yeon-Min I., Hisashi D., Naoyuki N., Takao H. Surface structures and osteoblast response of hydrothermally produced CaTiO3 thin film on Ti–13Nb–13Zr alloy. *Applied Surface Science*, volume 257, 2011: 7856-7863.
- Porter A. E., Punam T., Hobbs-Linn W., Coathup M.J., Gordon W.B., Spector M. Bone bonding to hydroxyapatite and titanium surfaces on femoral stems retrieved from human subjects at autopsy. *Biomaterials*, volume 25, 2004: 5199-5208.
- Rho Y.J., Kuhn-Spearing L., Zioupos P. Mechanical properties and the hierarchical structure of bone. *Medical Engineering & Phy*sics, volume 20, 1998: 92-102.

- Rokkum M., Reigstad A. Total hip replacement with an entirely hydroxyapatite-coated prosthesis: 5 years' follow-up of 94 consecutive hips. *The Journal of Arthroplasty*, volume 14 (issue 6), September, 1999: 689-700.
- Roushown A. and Masatomo Y. Space group and crystal structure of the Perovskite CaTiO3 from 296 to 1720K. *Journal of Solid State Chemistry*, volume 178, 2005: 2867-2872.
- Tang Y.C., Katsuma S., Fujimoto S., Hiromoto S. Electrochemical study of Type 304 and 316L stainless steels in simulated body fluids and cell cultures. *Acta Biomaterialia*, volume 2, 2006: 709-715.
- Wang W., Mingrong S., Li Y., Liang F., Fengang Z., Xinglong W. Enhanced photoluminescence of CaTiO3:Pr3+ phosphor films deposited on SiO2 buffered Si substrates. *Thin Solid Films*, volume 517, 2009: 3398-3401.
- Wiff J.P., Fuenzalida V.M., Arias J.L., Fernandez M.S.. Hydrothermal–electrochemical CaTiO3 coatings as precursor of a biomimetic calcium phosphate layer. *Materials Letters*, volume 61, 2007: 2739-2743.
- Yang H.K., Jong-Won C., G.-Seeta R.R., Byung-Kee M., Byung-Chun C., Jung-Hyun J., Jung-Hwan K. Luminescent characteristics of CaTiO3:Pr3+ thin films prepared by pulsed laser

deposition method with various substrates. *Applied Surface Science*: volume 255, 2009: 5062-5066.

Zhanga F., Maa A., Jianga J., Xua H., Songa D., Lua F., Nishida Y. Enhanced biodegradation behavior of ultrafine-grained ZE41A magnesium alloy in Hank's solution. Progress in Natural Science. *Materials International*, volume 23 (issue 4), 2013: 420-424.

Citation for this article:

Chicago style citation

Esguerra-Arce, Johanna, Yesid Aguilar-Castro, William Aperador-Chaparro, Leonid Ipaz-Cuastumal, Gilberto Bolaños-Pantoja, Carlos Alberto Rincón-López. Tribological Behavior of Bone Against Calcium Titanate Coating in Simulated Body Fluid. *Ingeniería Investigación y Tecnología*, XVI, 02 (2015): 279-286.

ISO 690 citation style:

Esguerra-Arce J., Aguilar-Castro Y., Aperador-Chaparro W., Ipaz-Cuastumal L., Bolaños-Pantoja G.B., Rincón-López C.A. Tribological Behavior of Bone Against Calcium Titanate Coating in Simulated Body Fluid. *Ingeniería Investigación y Tecnología*, volume XVI (issue 2), April-June 2015: 279-286.

About the authors

- *Johanna Esguerra-Arce*. She obtained his bachelor's degree in Materials Engineering at The Universidad del Valle, Colombia, and is a Ph.D. candidate in Engineering Doctorate with major in Materials Engineering. Her work has been presented in various national and international forums and is aimed at biomaterials and coatings obtained by physical vapor deposition methods. She has been occasional professor in Department of Material Engineering at The Universidad del Valle.
- *Yesid Aguilar-Castro*. He holds a master's in Metallurgy and Materials Science and a Ph. D in Materials and its New Technologies from Universidad Politécnica de Valencia. He has more than twelve years of professional experience as a professor at the Universidad del Valle and is appointed as a full time professor at the school of Materials Engineering. He has dedicated his work to tribology and supervises thesis work at the bachelor, master and doctoral levels.
- *William Aperador-Chaparro*. Research Professor of Universidad Militar Nueva Granada, PhD in Materials Engineering from the Universidad del Valle. He obtained the degree of Master of Metallurgy and Materials Science from Universidad Pedagógica y Tecnológica de Colombia (UPTC-Tunja), he also has a degree in physicist from the same university. He is a researcher of the Volta group in Mechatronics Engineering program.
- *Leonid Ipaz-Cuastumal.* Professor at Universidad Autónoma de Occidente. PhD in Materials Engineering from the Universidad del Valle, also has a degree in Materials Engineer from the same university. His main research interests are synthesis and characterization of hard coatings. He is a researcher of the Thin Films Group in the Physics Department of The Universidad del Valle.
- *Gilberto Bolanos-Pantoja*. He obtained his bachelor's degree in physics at the Universidad del Valle, Master and Doctoral studies of science (physics) at the same university. He is a research professor at The Universidad del Cauca. The research is aimed at manufacturing of multilayer hard coatings and phase transitions. He has directed 14 B.Ing. theses, and co-directed 2 doctoral theses.
- *Carlos Alberto Rincón-López.* His doctoral degree is from The Universidad del Valle (science-physics) and holds a master in physics from Universidad Simón Bolívar. At present he is an associate professor at The Universidad del Cauca. His research work has centered on plasma physics and coatings obtained by physical vapor deposition. He has directed various B.Ing. theses and has participate as a referee in 10 master and doctoral theses.