INGENIERÍA INVESTIGACIÓN Y TECNOLOGÍA volumen XXI (número 3), julio-septiembre 2020 1-13 ISSN 2594-0732 FI-UNAM artículo arbitrado Información del artículo: Recibido: 27 de octubre de 2018, reevaluado: 5 de agosto de 2019, aceptado: 10 de junio de 2020 Attribution-NonCommercial-NoDerivatives 4.0 International (CC BY-NC-ND 4.0) license https://doi.org/10.22201/fi.25940732e.2020.21.3.025



In-Silico study of microwave ablation applicators of different size for breast cancer treatment

Estudio de aplicadores de ablación por microondas de diferentes dimensiones para el tratamiento del cáncer de mama

Hernández-Jácquez José Irving Instituto Tecnológico de la Laguna, Torreón Tecnológico Nacional de México División de Estudios de Posgrado e Investigación E-mail: joseirving81@gmail.com https://orcid.org/0000-0002-5590-9735

Cepeda-Rubio Mario Francisco Jesús Instituto Tecnológico de la Laguna, Torreón Tecnológico Nacional de México División de Estudios de Posgrado e Investigación E-mail: karmario@gmail.com https://orcid.org/0000-0003-0310-0712

Guerrero-López Geshel David Instituto Tecnológico de la Laguna, Torreón Tecnológico Nacional de México División de Estudios de Posgrado e Investigación E-mail: geshel87@gmail.com https://orcid.org/0000-0003-0341-501X

Vera-Hernández Arturo Instituto Politécnico Nacional, México Centro de Investigación y de Estudios Avanzados Sección de Bioelectrónica E-mail: arvera@cinvestav.mx

https://orcid.org/0000-0001-6258-154X

Leija-Salas Lorenzo Instituto Politécnico Nacional, México Centro de Investigación y de Estudios Avanzados Sección de Bioelectrónica E-mail: lleija@cinvestav.mx https://orcid.org/0000-0001-8437-6520

Valdés-Perezgasga Francisco Instituto Tecnológico de la Laguna, Torreón Tecnológico Nacional de México División de Estudios de Posgrado e Investigación E-mail: fvaldesp@gmail.com https://orcid.org/0000-0002-7176-3641

Flores-García Francisco Instituto Tecnológico de la Laguna, Torreón Tecnológico Nacional de México División de Estudios de Posgrado e Investigación E-mail: francisco.floresgarcia@gmail.com https://orcid.org/0000-0002-6661-8191

Abstract

Breast cancer is one of the most common types of cancer in females around the world, surgery is the preferred treatment for this disease in early stages, however there are some emergent treatment options with less clinical and cosmetic repercussions, such as the microwave ablation. In order to determine an appropriate design for applicators that could be used for the treatment of breast cancer, an in-silico study of microwave ablation of breast cancer was conducted to show the difference between micro-coaxial applicators constructed with two different types of coaxial cable (UT-47 and UT-85) utilizing the same design in three possible scenarios: breast tissue, tumor tissue and tumor surrounded by breast tissue. The simulation was performed using COMSOL Multiphysics commercial software, due to its capabilities to simulate the electromagnetic and thermals effects. A slot type applicator was selected because of its ease of manufacture and the available literature on it. The results show a better coupling for the applicator constructed with the UT-47 cable, especially in the phantom tumor surrounded by breast phantom test. Nevertheless, is necessary to validate these results in ex-vivo breast cancer tissue.

Keywords: Microwave, ablation, SWR, applicator, computer simulation.

Resumen

El cáncer de mama es uno de los tipos más comunes de cáncer en las mujeres alrededor del mundo, la cirugía es el tratamiento más común para esta enfermedad en etapas tempranas, sin embargo ,existen algunos métodos emergentes como opciones para tratar esta enfermedad con menores repercusiones clínicas y cosméticas, donde una de estas es la ablación por microondas. Para determinar un diseño apropiado del aplicador para emplear el tratamiento de cáncer de mama, se realizó un modelo in-silico de la ablación por microondas de cáncer de mama, para mostrar las diferencias entre aplicadores microcoaxiales construidos con dos tipos de cable coaxial (UT-47 y UT-85), utilizando el mismo diseño en tres posibles escenarios; tejido mamario, tejido tumoral, y tejido tumoral rodeado por tejido mamario. La simulación se realizó utilizando el software comercial COMSOL Multiphysics, debido a sus capacidades de simular los efectos electromagnéticos y térmicos. Se seleccionó un aplicador de tipo ranura dada su fácil manufactura y la cantidad de literatura disponible sobre él. Los resultados muestran un mejor acoplamiento para el aplicador construido con el cable UT-47, especialmente en la prueba realizada en phantom tumoral rodeado por phamtom mamario. Sin embargo, es necesario validar estos resultados en tejido ex vivo de cáncer de mama.

Descriptores: Ablación, microondas, SWR, aplicador, modelo computacional.

INTRODUCTION

A tumor is a mass of transformed cells with abnormal growth and multiplication (The Oxford English Dictionary, 2010). Tumors can be cancerous (malignant) or non-cancerous (benign) and occur when cells multiply excessively in the body. Normally, cell division and growth are tightly controlled. New cells are created to replace old ones. Cells that are damaged or are no longer needed, die to make way for healthy replacement cells. If the balance of division and cell death is altered, a tumor may form (U.S. National Library of Medicine, 2014).

Breast cancer appears as malignant tumors originating in the cells of the breast. Breast cancer occurs almost exclusively in women, but cases have been known to occur in men. Breast cancer is the second most common cancer worldwide with 1.677 million cases equivalent to 11.9 % of all cancers (Who. GLOBOCAN, 2012). In developed countries, the survival rate at 5 years for this type of cancer exceeds 80 %, while for underdeveloped countries the rate falls between 10 to 40 %, this is because developing countries often lack the facilities and personnel for an early detection of the disease (WHO, 2014).

Advances in science have yielded new alternatives for treating this type of cancer. Often, the patient has several options: chemotherapy, radiotherapy, hormone therapy and surgery, the latter is the most widely used in the treatment of breast cancer (Breast Cancer Org., 2019).

If the cancer is detected at an early stage, there are minimally invasive techniques for treatment, for example: laser photocoagulation, ultrasound, cryotherapy and hyperthermia ablation; which is divided into radiofrequency ablation (RFA) and microwave ablation (MWA) (Van *et al.*, 2007). Additionally, these minimally invasive techniques can be applied to patients who for various reasons cannot be treated surgically (Hompes *et al.*, 2010; Zanus *et al.*, 2011; Simo *et al.*, 2013; Lloyd *et al.*, 2011; Ratanaprasatporn *et al.*, 2013; Livraghi *et al.*, 2012; Iannitti *et al.*, 2007; Veltri *et al.*, 2012; Carrafiello *et al.*, 2014; Little *et al.*, 2013; Wolf *et al.*, 2012; Carrafiello *et al.*, 2010; Guan *et al.*, 2012).

The placement of a needle or a catheter directly into a tumor and the use of heat, cold, or a chemical to destroy it, is called ablation. It is more frequently used to stop the spread of cancer to the bones or the liver, although it can also be used in other places or organs. Ablation is usually employed when only a limited number of tumors are causing problems.

A common type of ablation, radiofrequency ablation (RFA), uses a needle that carries an electric current. The end of the needle is inserted into the tumor. An ultrasound or a CT (Tomography) can be used to ensure that the needle is in the right place. An electrical current that passes through the needle warms the tumor to destroy it. In general, RFA is performed while the patient is under general anesthesia (Vydra & Vrba, 2011).

Another type of ablation, called Cryoablation, uses the probe placed in the tumor to freeze it, which destroys cancer cells (Shinohara, 2007).

Other methods use heat (laser-induced interstitial thermotherapy) or alcohol to destroy cells (Society AC, 2015).

In medicine, microwaves are being used increasingly for tumor ablation. Microwaves cause the tumor to heat and burn. To achieve MWA the tissue should be heated to a temperature of 55 °C or more, however maximum temperature should not reach 100 °C, since that would cause water to evaporate changing tissue properties, several prototypes of antennas are used to accomplish that. These antennas are designed trying to make them smaller and better targeting of energy to prevent damage on healthy tissue (Bertram *et al.*, 2006).

MWA is a promising technique for breast cancer because the energy can be focused to selectively damage carcinomas of the breast. This can be attained due to their high water content compared with a lower level of heating in the adipose and glandular tissue that have lower water content (Vargas *et al..*, 2004; Mariya *et al..*, 2007; Liu *et al..*, 2018).

For breast cancer, reports of the application of this technique have been made in patients with inoperable tumors (Zhou *et al..*, 2014), the results obtained so far are an excellent option for people suffering from this disease. The results have shown that it is a very good alternative in small tumors, since the application time is small and the results are very encouraging especially for these patients (Liu *et al..*, 2018; Li *et al..*, 2017).

There are several prototypes of antennas that have been reported with their associated simulations and ex vivo tissue studies, mostly in swine breast or bovine liver (Guerrero, 2014; Cavagnaro *et al..*, 2011; McWilliams *et al..*, 2015; Luyen *et al..*, 2015; Karampatzakis *et al..*, 2013), in the Table 1 a comparison of different types of antennas and their results is shown.

The closed tip antenna has an outer diameter of 1.53 mm however its construction is complicate, the directional antenna has 1.194 mm outer diameter but it needs a reflector that has a diameter of 2.4 mm, the helical antenna has an outer diameter of 2.197 mm, finally the slot and double slot antennas were constructed with a diameter of 1.3 mm. Every mentioned antenna show good results, however the slot and double slot antennas are easier to construct, for these reasons a slot antenna was considered for this study.

Table 1. Companion of difference (ypes of alternitia and their results)					
Reference	Antenna type	Experiment	Results		
Cavagnaro, et al., 2011	Closed tip antenna	Bovine <i>ex-vivo</i> liver	Ablation zone of approx. 38 x 28 mm (Length × Diameter) with a 20 watts		
McWilliams, et. al., 2015	Directional antenna	Porcine muscle	potency Ablation radius 10 mm in forward direction and 5 mm radius in backwards direction, with a 80 watts power		
Luyen, et. al., 2015	Helical antenna	Bovine <i>ex-vivo</i> liver	Ablation zone of 75 x 43 mm, with a 42 watts power.		
Karampatzakis, et. al., 2013	Slot-antenna Double slot-antenna	In-silico study	Ablation zone of 32 x 30.5 mm, with a 50 watts power		

Table 1	Comparison	of different tv	oes of antennas	and their results
---------	------------	-----------------	-----------------	-------------------

The objective of this work is to evaluate the efficiency of two applicators with the same design, slot antenna (Cepeda *et al.*, 2015), but different diameters, to determine the feasibility of using a thinner applicator, which facilitates puncture, since it is important to select the most suitable type of cable for microwave ablation.

MATERIALS AND METHODS

SLOT ANTENNA

MWA applicators are antennas, generally coated with polytetrafluoroethylene (PTFE) to minimize adhesion to the (carbonized) tissue that is exposed to ablation, normally, the applicators designs for MWA are made for treatment of liver cancer, and some of these designs can be: Dipole, Monopole and Slot.

The design chosen for the present work is a slot antenna. In MWA, the slot antenna is one of the most popular designs, it is easily made with a semi rigid coaxial cable, which is short-circuited at one end and a ringshaped slot is cut in the outer conductor of the cable (Cepeda *et al.*, 2015), a diagram of this design is shown in the Figure 1. In this work two slots antennas were considered, one with a diameter of 2.197 mm and the second with a diameter of 1.194.

The effective wavelength at the operation frequency in the tumor tissue was considered to select the distance between the tip of the antenna and the slot. To calculate the wavelength the equation (1) is used:



In the equation *c* represents the speed of light in free space (m/s), *f* represents the operating frequency of the generator (2.45 GHz), ε_r represents relative permittivity of the tumor at operating frequency (59.385) and μ_r represents the magnetic relative permeability (1). Using this values the wavelength obtained is 15.88 mm; but since the tissue is not homogeneous, this value can be used only as a reference. The distance between the slot and the tip is approximately one fourth of the effective wavelength.

Standing Wave Ratio

The standing wave ratio (SWR) is the ratio of maximum to minimum voltage signal on a transmission line (Merriam-Webster, 2019). Ideally, a transmission line should have a SWR of 1, with all the power delivered and no power reflected. To calculate the SWR the equation (2) is used:

$$SWR = \frac{1+|\Gamma|}{1-|\Gamma|} \tag{2}$$

Where Γ represents the reflection coefficient. Before starting the ablation the SWR should be measured and during the ablation test the incident and reflected pow-



Figure 1. Diagram the slot antenna used for the simulation

ers must be monitored. This is because there is a warming called detrimental heating, which occurs along the antenna when a large portion of the energy is reflected to the signal generator instead of being delivered to the tissue. This heating can cause damage in tissue outside the treatment region, and if the SWR is too high it can damage the equipment. To measure SWR a network analyzer (E5071B ENA, Agilent Technologies, Santa Clara CA) was used, this device was also used in the measurement of the dielectric properties of the materials used for validation experiments.

IN SILICO MODEL

Simulations were made to observe the performance of antennas before doing experimental validation in phantom tissue. The thermal and dielectric properties of breast cancerous tumors and healthy breast tissue were obtained from the literature (Mariya *et al.*, 2007; Lazebnik et al., 2007; Gautherie, 1980), the values used in the computer model are shown in Table 2.

Table 2. Dielectric properties of breast tumor tissue at 2.45 GHz

Property	Breast Tissue	Cancerous Tissue
Conductivity (s/m)	0.137	3.156
Relative Permittivity	5.1467	59.385
Thermal Conductivity (W·m ⁻¹ ·K ⁻¹)	0.42	0.50

The computational model was made in COMSOL Multiphysics (COMSOL Inc., Burlington MA), due to its

35 Breast tissue 30 25 Applicator 20 15 10 Slot 5 0 Breast carcinoma mm 25 35 40 45 50 10 15 20 30

ability to simulate the electromagnetic and thermal characteristics of the problem. COMSOL Multiphysics uses the finite element method (FEM) to solve the physical problems. The FEM is a numerical technique that can be formulated as functional minimization, it involves dividing a complex geometry into small elements for a system of partial differential equations, evaluated at nodes or edges. Nonetheless to obtain an accurate solution, appropriate initial and boundary conditions should be selected.

The simulation can be reduced to a 2D axisymmetric model, since it presents rotational symmetry, allowing a two-dimensional modeling using cylindrical coordinates, and reducing computational time. Figure 2 shows the in silico model considering an applicator inserted in a round breast carcinoma surrounded by healthy breast tissue.

The considered boundary conditions are: for both simulations the z axis is considered as axial symmetry, for the electromagnetic simulation all the exterior boundaries are considered as scattering boundary condition, with the only exception of the exterior boundary of the dielectric of the applicator, which is considered as port condition with and applied power of 10 W, for thermal simulation all the exterior boundaries are considered as constant temperature (25 °C). As mentioned above the FEM divides the geometry into smaller elements, this process is called meshing. The accuracy that can be obtained from any finite element analysis (FEA) depends on the used mesh. As mesh elements are made smaller the computed solution will be closer to the actual solu-

Figure 2. Model geometry considering a UT-47 applicator inserted in a round breast carcinoma surrounded by healthy breast tissue

tion. However, if the mesh is excessively thin the computational time and resources needed to obtained the solution are higher. Taking this into consideration the mesh should be thin enough to obtain a valid solution, but not too much to make computational time tend to infinite. For an electromagnetic FEA the maximum element size recommended is one eighth of the effective wavelength at highest frequency. Using equation (1), considering that the maximum frequency in the study is 3.5 GHz and the highest relative permittivity is the one for the breast carcinoma (ε_r =59.385), the smallest wavelength in the study is 11.11 mm, so the maximum element size shouldn't exceed 1.38 mm. The mesh used in the computational model has a maximum element size of 1 mm and a minimum size of 0.0038 mm, under these conditions the mesh has 38638 triangular elements, 2528 edge elements, 23 vertex elements. A portion is shown in Figure 3 of the mesh used in the model.

For computational modeling it is necessary to use an equation that governs the physical behavior of the phenomenon, for the case of ablation therapies the most used equation is the Pennes bioheat equation (Ibitoye *et al.*, 2015) that is given by:

$$\rho C_p \frac{dT}{dT} + \nabla \cdot (-K\nabla T) = \rho_{bl} C_{bl} \omega_{bl} (T_{bl} - T) + Q_{met} + Q_{ext}$$
(3)

In the equation ρ is the tissue density (kg·m⁻³), C_p is the specific heat capacity at constant pressure (J·kg·m⁻³) k represents tissue thermal conductivity (W·m⁻¹·K⁻¹), ρ_{bl} is

the blood density (kg·m⁻³), C_{bl} represents the specific heat capacity of the blood (J·kg⁻¹·K⁻¹), ω_{bl} is the blood perfusion (kg·m⁻³·s⁻¹), T_{bl} represents blood temperature (K) and *T* is the final temperature (K), Q_{met} is the metabolic heat generation (W·m⁻³), Q_{ext} is the absorbed electromagnetic energy (W·m⁻³). Since the experimental validation is conducted on phantom tissue blood perfusion and metabolic heat generation can be neglected. The absorbed electromagnetic energy is given by the equation:

$$Q_{ext} = \frac{1}{2}\sigma |E|^2 \tag{4}$$

EXPERIMENTAL VALIDATION

ANTENNA CONSTRUCTION

Two applicators were constructed to validate the computational model with a UT-85 and UT-47 (Li *et al.*, 2017; Guerrero, 2014) coaxial cables, with 2.197 mm and 1.194 mm diameters respectively. The two applicators have the design shown in Figure 1. To build the applicators PCB manufacturing methods are used, first a mask is placed in the coaxial cable, after that the cable is painted with acrylic paint to protect the copper from the ferric chloride (Figure 4). Once the slots are made, the outer and inner conductor are short-circuited at the



Figure 3. Portion of the mesh used for the FEA, the finest mesh is in the slot of the antenna

tip of the applicator. Finally, a SMA connector is mounted on the other side of the applicator.

PHANTOM ELABORATION

A phantom is a physical model reproducing the electrical characteristics of interest, such as permittivity and conductivity. A phantom model should imitate dielectric properties of living tissue. Phantoms are usually made using sodium chloride (NaCl), aqueous solutions, glycerin, agar, aniline, among others, made based on NaCl solution, the most popular phantom. This is easily prepared and its components are readily and easily available. Distilled water is used and NaCl is added to adjust the conductivity without changing the permittivity with concentrations below 10 g / l. To adjust the permittivity vegetable oil is used and to achieve a homogeneous mix neutral detergent is added, finally to obtain a gelatinized phantom, agar is added, to the mix (Ortega, 2010).

To elaborate the breast phantom tridistilled water (50 ml), vegetable oil (150 ml), neutral detergent (30 ml) and agarose (4.5 g) were mixed and heated, the breast carcinoma phantom was composed of tridistilled water (100 ml), ethanol (60 ml), sodium chloride (1 g) and agarose (1.5 g) (Ortega, 2010).

The geometry utilized in the in-silico model is a spherical cancerous tumor inside a cylinder of breast tissue, to achieve this a sphere of cancerous phantom was created and once it solidified, it was inserted in a glass beaker filled with breast phantom not yet solidified, Figure 5.

SWR MEASUREMENT

To evaluate the efficiency of the applicators (constructed with 2.2 mm and 1.2 mm diameters cables), the SWR was measured with a network analyzer (E5071B ENA, Agilent Technologies, Santa Clara CA), when the applicator was inserted on the phantom. This measure was later compared with the SWR obtained in the in-silico model.

RESULTS

SWR

Three conditions were considered for the in-silico model and measurement of the SWR, when the applicator is inserted in healthy breast tissue when it is inserted into breast tumor tissue and when is inserted into a spherical tumor surrounded by healthy breast tissue.

For the first consideration (applicator inserted in healthy breast tissue) the SWR for the applicator constructed with a UT-47 microcoaxial cable (1.2 mm diameter) is 1.72 in the experimental validation while the in-silico model results show a SWR of 2.18, both of them at the operation frequency of 2.45 GHz, Figure 6 shows the SWR graph in the 2 to 3.5 GHz interval.

Considering the applicator constructed with a UT-85 microcoaxial cable (2.2 mm diameter) inserted in healthy breast tissue the SWR measured in the valida-



Figure 4. Process of fabrication of the applicators



Figure 5. Insertion of cancerous phantom in the liquid breast phantom

tion experiment was 1.50 while the in-silico value was 1.68 both for a 2.45 GHz frequency. Figure 7 shows the SWR graph in the 2 to 3.5 GHz interval.

For the second consideration (applicator inserted into breast tumor phantom), the SWR obtain in the validation experiment using the UT-47 applicator was 1.61, while the *in-silico* result was 1.18. The Figure 8

shows the SWR graph for a frequency of 2 GHz to 3.5 GHz.

For the second consideration (applicator inserted into breast tumor phantom), the SWR obtain in the validation experiment using the UT-85 applicator was 1.23, while the in-silico result was 1.54. The Figure 9 shows the SWR graph for a frequency of 2 GHz to 3.5 GHz.



Figure 6. Measured (solid line) and simulated (dashed line) SWR for the UT-47 microcoaxial cable applicator into breast phantom tissue, in the 2 to 3.5 GHz interval

Figure 7. Measured (solid line) and simulated (dashed line) SWR for the UT-85 microcoaxial cable applicator into breast phantom tissue, in the 2 to 3.5 GHz interval



Finally when the UT-47 applicator was inserted in a spherical tumor phantom surrounded by breast phantom, the SWR at a frequency of 2.45 GHz for the in-silico model was 1.17 while the measured in the validation experiment was 1.35. Figure 10 shows the SWR graph in the 2 to 3.5 GHz interval.

And for the applicator constructed with the UT-85 inserted in a spherical tumor phantom surrounded by breast phantom, the SWR at a frequency of 2.45 GHz for the in-silico model was 1.63 and the measured in the validation experiment was the same. Figure 11 shows the SWR graph in the 2 to 3.5 GHz interval.

Finally a comparative of the SWR measured for both applicators is shown in Figure 12, the minimum value of SWR for the applicator constructed with UT-47 coaxial cable is 1.14 at a frequency of 2.58 GHz, and the minimum value for the UT-85 applicator is 1.52 at a frequency of 2.91 GHz.

HEATING PATTERNS

The heat patrons for the three proposed conditions were only obtained in the in-silico model considering an applied power of 10 W for the port boundary of the applicator, for the first condition, applicator inserted into breast phantom, the maximum temperature achieved by the UT-47 applicator was 118 °C, the maximum ablation radius was 9.5 mm. Figure 13 shows the heating pattern for these conditions. Notice that maximum temperature was above 100 °C, however since the effect of blood perfusion was not considered in the simulation, it is expected that the maximum temperature reduces when the blood perfusion effect is present.

Similarly Figure 14 shows the heating pattern for the in-silico model considering the UT-85 applicator inserted into breast tissue, notice that the maximum temperature reached was lower (101 °C) for this applicator, and also the 55 °C isothermal contour was slightly smaller (0.1 mm in diameter and 5 mm in height).

For the second consideration (applicator inserted in tumor tissue), the heating pattern for the UT-47 applicator is shown in Figure 15, the maximum ablation radius was 11.7 mm and the temperature reached was 155 °C, the shape of the lesion was more spherical than the one obtained with the applicator inserted into breast tissue.



Figure 9. Measured (solid line) and simulated (dashed line) SWR for the UT-85 microcoaxial cable applicator into breast cancer phantom tissue, in the 2 to 3.5 GHz interval





Figure 11. Measured (solid line) and simulated (dashed line) SWR for the UT-85 microcoaxial cable applicator in tumor phantom surrounded by breast phantom, in the 2 to 3.5 GHz interval

Figure 12. Comparative of the SWR for applicators constructed with UT-85 (solid) and UT-47 (dashed). Notice that UT-47 applicator as a reflected power of less than 5% for almost all band

Figure 13. Heating pattern obtained in the simulation considering the UT-47 applicator inserted into breast tissue

Whereas the maximum temperature obtained conside-
ring the UT-85 applicator inserted in tumor tissue was
124 °C and the maximum ablation radius was 11.1 mm,
the heating pattern is shown in Figure 16.by breast tissue is show
temperature reached v
tion radius of 12.9 m
was not achieved.

In addition, the heating pattern of UT-47 applicator when it was inserted in a spherical tumor surrounded by breast tissue is shown in Figure 17, the maximum temperature reached was 161 °C with a maximum ablation radius of 12.9 mm, however total tumor ablation was not achieved.

Finally, the heating pattern of UT-85 applicator when it was inserted in a spherical tumor surrounded



the simulation considering the UT-85

the simulation considering the UT-47

the simulation considering the UT-85 applicator inserted in tumor tissue

by breast tissue is shown in Figure 18, the maximum temperature was 128 °C, and the maximum ablation radius was 12.3 mm.

In Table 3 the SWR, maximum ablation radius and maximum temperature for both applicators in the three experimental setup are shown. Notice that for the clinical situation, tumor surrounded by breast tissue, the results for the UT-47 applicator are better, since it has a superior coupling, a bigger ablation radius and a higher maximum temperature.

Hernández-Jácquez J.I., Cepeda-Rubio M. F. J., Guerrero-López G.D., Vera-Hernández A., Leija-Salas L, Valdés-Perezgasga F., Flores-García F.





Figure 18. Heating pattern obtained in the simulation considering the UT-85 applicator inserted in a spherical tumor surrounded by breast tissue

Table 3. SWR, maximum ablation radius and maximum temperature for UT-85 and UT-47 applicators in breast tissue, cancer tissue and cancer surrounded by breast tissue

		UT-85	UT-47
Brest Tissue	SWR (<i>In silico</i> /validation)	1.68/1.5	2.18/1.72
	Ablation Radius (mm)	9.4	9.5
	Max temperature (°C)	101	118
Cancer Tissue	SWR (<i>In silico</i> /validation)	1.54/1.23	1.18/1.61
	Ablation Radius (mm)	11.1	11.7
	Max temperature (°C)	124	155
Breast-Cancer Tissue	SWR (<i>In silico</i> /validation)	1.63/1.63	1.17/1.35
	Ablation Radius (mm)	12.3	12.9
	Max temperature (°C)	128	161

CONCLUSIONS

As an alternative to surgery, microwave ablation therapy is a promising technique to treat breast cancer in an early stage. The purpose of this work was to determine the viability of building thinner applicators, in order to reduce the size of the incision necessary to use this technique. The analyzed results were: temperature reached, size of lesion, and SWR, of in-silico and physical models (tumor and breast phantom).

The results obtained in the in-silico model show that the ablation temperatures are reached, in other words, temperature higher than 55 °C, it is important to notice that in all three considerations the temperature reached by the UT-47 applicator was superior to the obtained with the UT-85 applicator, however this is not necessarily good since temperatures above 100 °C can change tissue properties drastically, thus decoupling the applicator. Nonetheless since blood perfusion was not considered in the simulation these temperatures may be lower in a clinical situation. The coupling of the designed antennas was predicted using computer simulation. In general, both applicators show good coupling (SWR lower than 1.8). UT-47 applicator had higher value of SWR in the model where it was inserted into breast tissue compared to the SWR obtained by the UT-85 applicator, 2.18 and 1.68 respectively, but the UT-47 applicator has lower SWR in the other two computational models. The difference between simulations an the validation experiments does not present a problems since the reflected power for the UT-47 in the simulation in which the antenna is inserted in tumor tissue surrounded by breast tissue was 0.6 % while in the validation experiment was 2.2 %, on the other hand the results for the antenna constructed with the UT-85 in the same condition were the identical in the simulation and the validation experiment, 5.7 % of reflected power, this lead us to believe that the computational model is appropriated to predict the reflected power. The differences between the in-silico model and the validation experiments could be explained by the fact that the dielectric properties of tissue were not as constant and homogenous as they were considered for the simulation. It is also important to notice that in the simulations where the applicators were inserted into a spherical tumor surrounded by breast tissue, total ablation was not achieved, though it may be obtained by increasing the applied power.

Unquestionably the mechanical properties of the applicators should be considered, and the thinner applicator (UT-47) is extremely fragile, since it presented difficulties to be inserted in the phantoms, because it tended to bend easily. While the UT-85 applicator did not presented this problem And this must be considered with great importance when selecting the most suitable cable in microwave ablation.

Lastly it is still necessary to validate these results in ex-vivo breast cancer tissue and in-vivo laboratory tests before concluding if it is feasible to use these applicators for breast cancer treatment.

REFERENCES

- Bertram, J. M., Yang, D., Converse, M. C., Webster, J. G. & Mahvi, D. M. (2006). A review of coaxial-based interstitial antennas for hepatic microwave ablation. *Critical Reviews in Biomedical Engineering*, 34(3), 187-213. https://doi.org/10.1615/critrevbiomedeng.v34.i3.10
- Cancer B. treatment & side effects (2019). Retrieved on October from.www.breastcancer.org/treatment
- Carrafiello, G., Mangini, M. De Bernardi, I. *et al.* (2010a). Microwave ablation therapy for treating primary and secondary lung tumours: technical note. *La radiología médica*, 115(6), 962-974. https://doi.org/10.1007/s11547-010-0547-7
- Carrafiello, G., Mangini, M., Fontana, F. *et al.* (2014). Microwave ablation of lung tumours: single-centre preliminary experience. *La radiología médica*, 119(1), 75-82. https://doi.org/10.1007/ s11547-013-0301-z
- Carrafiello, G., Mangini, M., Fontana, F. et al. (2010b). Single-Antenna microwave ablation under contrast-enhanced ultrasound guidance for treatment of small renal cell carcinoma: Preliminary Experience. CardioVascular and Interventional Radiology, 33(2), 367-374.
- Cavagnaro, M., Amabile, C., Bernardi, P., Pisa, S. & Tosoratti, N. (2011). A minimally invasive antenna for microwave ablation therapies: Design, performances, and experimental assessment. *Biomedical Engineering*, IEEE Transactions on, 58(4), 949-959. https://doi.org/10.1109/TBME.2010.2099657
- Cepeda-Rubio, M. F. J., Guerrero, G. D., Valdés, F., Flores, F., Vera, A. & Leija, L. (2015). Computer modeling for microwave ablation in breast cancer using a coaxial slot antenna. *International Journal of Thermophysics*, 36(10), 2687-2704. https://ui.adsabs. harvard.edu/link_gateway/2015IJT....36.2687C/doi:10.1007/ s10765-015-1931-2
- Gautherie, M. (1980). Thermopathology of breast cancer: Measurement and analysis of in vivo temperature and blood flow. *Annals of the New York Academy of Sciences*, 335(1), 383-415. https://doi.org/10.1111/j.1749-6632.1980.tb50764.x
- Guan W., Bai J., Liu J. et al. (2012). Microwave ablation versus partial nephrectomy for small renal tumors: Intermediate-term results. Journal of Surgical Oncology, 106(3), 316-321. https:// doi.org/10.1002/jso.23071
- Guerrero, G. D. (2014). Desarrollo de una antena de ablación por microondas para su estudio en cáncer de mama. Torreón, Coahuila de Zaragoza, México: Departamento de Estudios de Posgrado e Investigación, Instituto Tecnológico de la Laguna.
- Hompes, R., Fieuws, S., Aerts, R., Thijs, M., Penninckx, F. & Topal, B. (2010). Results of single-probe microwave ablation of metastatic liver cancer. *European Journal of Surgical Oncology* (EJSO), 36(8), 725-730.
- Iannitti, D. A., Martin, R. C. G., Simon, C. J. *et al.* (2007). Hepatic tumor ablation with clustered microwave antennae: the US

Phase II Trial. HPB. The Official Journal of the International Hepato Pancreato Biliary Association, 9(2), 120-124.

- Ibitoye, Z. A., Nwoye, E. O., Aweda, M. A., Oremosu, A. A., Annunobi, C. C. & Akanmu, O. N. (2015). Optimization of dual slot antenna using floating metallic sleeve for microwave ablation. *Medical Engineering* and Physics, 37(4), 384-391. https://doi. org/10.1016/j.medengphy.2015.01.015
- Karampatzakis, A., Kühn, S., Tsanidis, G., Neufeld, E., Samaras, T. & Kuster, N. (2013). Antenna design and tissue parameters considerations for an improved modelling of microwave ablation in the liver. *Physics in Medicine and Biology*, 58(10), 3191-3206. https://doi.org/10.1088/0031-9155/58/10/3191
- Lazebnik, M., Okoniewski, M., Booske, J. H. & Hagness, S. C. (2007). Highly accurate debye models for normal and malignant breast tissue dielectric properties at microwave frequencies. *IEEE Microwave and Wireless Components Letters*, 17(12), 822-824. https://doi.org/10.1109/LMWC.2007.910465
- Li, L., Wang, W., Pan, H. et al. (2017). Microwave ablation combined with OK-432 induces Th1-type response and specific antitumor immunity in a murine model of breast cancer. *Journal of Translational Medicine*, 15(1), 23. https://dx.doi.org/10.1186%2Fs12967-017-1124-9
- Little, M., Chun, D., Boardman, P., Gleeson, F. & Anderson, E. (2013). Microwave ablation of pulmonary malignancies using a novel high-energy antenna system. *CardioVascular and Interventional Radiology*, 36(2), 460-465. https://doi.org/10.1007/ s00270-012-0465-2
- Liu, B., Wu, Z., Mo, H. *et al.* (2018). Safety and efficacy of microwave ablation for breast cancer thoracic metastases. *Cancer Manag Res.*, 10, 5685-5689. https://doi.org/10.2147/cmar. s176743
- Livraghi, T., Meloni, F., Solbiati, L. & Zanus, G. (2012). Complications of microwave ablation for liver tumors: results of a multicenter study. *CardioVascular and Interventional Radiology*, 35(4), 868-874.
- Lloyd, D. M., Lau, K. N., Welsh, F. *et al.* (2011). International multicentre prospective study on microwave ablation of liver tumours: preliminary results. HPB, 13(8), 579-585.
- Luyen, H., Hagness, S. C. & Behdad, N. A. (2015). Balun-Free helical antenna for minimally invasive microwave ablation. *IEEE Transactions on Antennas and Propagation*, 63(3), 959-965.
- Mariya, L., Dijana, P., Leah, M. et al. (2007). A large-scale study of the ultrawideband microwave dielectric properties of normal, benign and malignant breast tissues obtained from cancer surgeries. *Physics in Medicine and Biology*, 52(20), 6093.
- McWilliams, B. T., Schnell, E. E., Curto, S., Fahrbach, T. M. & Prakash, P. (2015). A directional interstitial antenna for microwave tissue ablation: Theoretical and experimental investigation. *IEEE Transactions on Biomedical Engineering*, 62 (9), 2144-2150. https://doi.org/10.1109/tbme.2015.2413672
- Merriam-Webster, Dictionary (2019). Retrieved from http://www. merriam-webster.com/dictionary/standing%20wave%20ratio

- Ortega, R., Leija, L., Vera, A. & Cepeda, M. F. J. (2010). Measurement of breast-tumor phantom dielectric properties for microwave breast cancer treatment evaluation. Paper presented at Electrical Engineering Computing Science and Automatic Control (CCE). In 7th International Conference, 8-10 Sept.
- Ratanaprasatporn, L., Charpentier, K. P., Resnick, M., Lu, S. & Dupuy, D. (2013). Intra-operative microwave ablation of liver malignancies with tumour permittivity feedback control: a prospective ablate and resect study. *HPB*, 15(12), 997-1001. https://doi.org/10.1111/hpb.12084
- Shinohara, K. (2007). Cryotherapy. International Journal of Clinical Oncology. 212(6), 416-426.
- Simo, K. A., Tsirline, V. B., Sindram, D. et al. (2013). Microwave ablation using 915-MHz and 2.45-GHz systems: what are the differences? HPB, 15(12), 991-996.
- Society A. C. breast cancer. (2015). Retrieved on October from http://www.cancer.org/cancer/breastcancer
- The Oxford English Dictionary (2010). In: Press OU.
- Vydra T. & Vrba J. (2011). Ablation applicator for destructive hyperthermia treatment. *Journal of Physics: Conference Series*, 329(1), 012035.
- U.S. National Library of Medicine. (2014). Retrieved from www. nlm.nih.gov
- Van, S., Van den Bosch, M. A. J., Van, P., Mali, W. M., Borel, I. M. & Van, R. (2007). Minimally invasive ablative therapies for invasive breast carcinomas: An overview of current literature. *World Journal of Surgery*, 31(12), 2284-2292.
- Vargas, H. I., Dooley, W. C., Gardner, R. A. et al. (2004). Focused microwave phased array thermotherapy for ablation of earlystage breast cancer: Results of thermal dose escalation. *Ann Surg Oncol.*, 11(2), 8. https://doi.org/10.1245/aso.2004.03.059
- Veltri, A., Gazzera, C., Rotondella, C., Camerano, F., Busso, M. & Gandini, G. (2012). Image-guided microwave ablation of hepatic tumours: preliminary experience. *La radiología*, 117(3), 378-392.
- WHO. GLOBOCAN (2012). Estimated cancer incidence, mortality and prevalence worldwide in 2012. Retrieved on 2013 from http://www.who.int/cancer/en/
- WHO. (2014). WHO position paper on mammography screening. 1a ed.
- Wolf, J., B.-A., Ng. & Damian, E. (2012). Dupuy. Intraoperative microwave ablation of pulmonary malignancies with tumor permittivity feedback control: Ablation and Resection Study in 10 Consecutive Patients. *Radiology*, 262(1), 353-360. https:// doi.org/10.1148/radiol.11110015
- Zanus, G., Boetto, R., Gringeri, E. et al. (2011). Microwave thermal ablation for hepatocarcinoma: Six liver transplantation cases. *Transplantation Proceedings*, 43(4), 1091-1094. https://doi. org/10.1016/j.transproceed.2011.02.044
- Zhou, W., Jiang, Y., Chen, L. *et al.* (2014). Image and pathological changes after microwave ablation of breast cancer: A pilot study. *European Journal of Radiology*, 83(10), 1771-1777. https://doi. org/10.1016/j.ejrad.2014.06.015